

计量授权证书附件

序号 No.	开展商品量/商品 包装计量检验的参 数名称 Parameters of Metrological Testing for Quantity/Package of Products	测量范围 Measurement Range	准确度等级或 测量扩展不确定度 Accuracy or Expanded Measurement Uncertainty	依据检定 规程编号 Number of Verification Specification Referred to
1	质量标注定量包装 商品净含量	(0~50) g (50~100) g (100~200) g (200~300) g (300~500) g 500g~1kg (1~10) kg (10~15) kg (15~50) kg	$\leq Q_n \times 1.8\%$ $\leq 0.9g$ $\leq Q_n \times 0.9\%$ $\leq 1.8g$ $\leq Q_n \times 0.6\%$ $\leq 3g$ $\leq Q_n \times 0.3\%$ $\leq 30g$ $\leq Q_n \times 0.2\%$	JJF 1070 JJF 1070.1 JJF 1070.2 JJF 1070.3
2	体积标注定量包装 商品净含量	(0~50) mL (50~100) mL (100~200) mL (200~300) mL (300~500) mL 500mL~1L (1~10) L (10~15) L (15~50) L	$\leq Q_n \times 1.8\%$ $\leq 0.9mL$ $\leq Q_n \times 0.9\%$ $\leq 1.8mL$ $\leq Q_n \times 0.6\%$ $\leq 3mL$ $\leq Q_n \times 0.3\%$ $\leq 30mL$ $\leq Q_n \times 0.2\%$	JJF 1070
3	长度标注定量包装 商品净含量	$Q_n \leq 5m$ $Q_n > 5m$	$\leq Q_n \times 0.2\%$ $\leq Q_n \times 0.4\%$	JJF 1070
4	计数标注定量包装 商品净含量	$Q_n \leq 50$ $Q_n > 50$	— $\leq Q_n \times 0.2\%$	JJF 1070
5	面积标注定量包装 商品净含量	全部 Q_n	$\leq Q_n \times 0.6\%$	JJF 1070

-----以下空白-----